

**University School of Basic and Applied Sciences
GGS Indraprastha University
Sector 16-C, Dawarka New Delhi
Phone No. 011-25302402**

GGSIU/USBAS/Store/2016-17/15285^L

January 13th, 2017

Sealed quotation are invited upon the approval of the competent authority of University School of Basic & Applied Sciences, Guru Gobind Singh Indraprastha University, Sector 16C, Dwarka, New Delhi-110078 from Manufacturer/ Distributors/ Authorized dealers for the supply of analytical balance (weighing balance). The technical specifications, terms and conditions are as follows:

Technical Specifications:

1	Maximum Capacity	220g
2	Maximum Capacity, fine range	82g
3.	Readability	0.01mg
4.	Repeatability (at nominal load)	0.08mg
5.	Linearity Deviation	0.2mg
6.	Linearity Deviation (within 10g)	0.02mg
7.	Minimum Sample Weight (acc. To USP)	45mg
8.	Minimum Sample Weight (U=1% K=2)	3mg
9.	Settling Time	4 s
10.	Settling time, fine range	8s
11.	Balance Dimensions (WxDxH in mm)	247x358x331
12.	Weighing pan dimensions (mm)	ø 80
13.	Fully Automated Calibration Technology	
14.	Display Type	Blacklit LCD

Terms and Conditions:

1. The firms must be manufacturer / authorized dealers / authorized distributor.
2. Payment: After satisfactory installation. No advance payment will be made.
3. Delivery within 30 days of placing the order.
4. Taxes: Tax as applicable may please be mentioned in the quotation.
5. Warranty: At least one year from the date of installation.
6. Quotation received after due date and time shall be summarily rejected.

Last date, time and venue for submission of quotation	Till 30.01.2017 up to 4 PM At Room No. 203, Block-B University School of Basic & Applied Sciences, Guru Gobind Singh Indraprastha University, Sector 16C, Dwarka, New Delhi -110078
---	---

Vaishali
13/1/2017
Prof. Vaishali Singh

Copy to:-

1. In-charge, Server Room – with a request to upload the same on University website.
2. Guard file