

University School of Chemical Technology GURU GOBIND SINGH INDRAPRASTHA UNIVERSITY

Sector: 16C, Dwarka, New Delhi-110 078

www.ipu.ac.in

Ref. No.: IPU/USCT/Pur/2014/

Date: 25.02.14

Sealed item rate quotations are invited along with the photograph of the following instrument on behalf of Dean, University School of Chemical Technology, Guru Gobind Singh Indraprastha University, Sector: 16C, Dwarka, New Delhi-110078 from manufacturers/distributors/authorized dealers for the supply of following item with the following technical specifications and terms and conditions:

Sr. No.	Name of the items	Technical Specifications
	Analytical Balance	Weighing capacity: 220 g
		Readability: 0.0001 g
		Repeatability: 0.0001 g
		Calibration: Internal
		Display Type: LCD
		Pan size: 90 mm diameter or more
		Balance should be covered with glass or transparent
		material during measurement.

Terms & Conditions:

- 1. The firms must be manufacturer/authorized dealer/authorized distributor.
- 2. FOR: GGS Indraprastha University, Sector: 16C, Dwarka, New Delhi-110078
 - 3. Payment: After satisfactory Installation. No advance payment will be made.
- 4. Delivery: 30 days
- 5. Taxes: Tax as applicable may please be clearly mentioned in the quotation.
- 6. Warranty: Three years
- 7. Quotations received after due date & time shall be summarily rejected.

Last date, time and venue for submission of quotation

O7.03.14

Maryly