



**Guru Gobind Singh Indraprastha University**  
**Sector 16-C, Dwarka, New Delhi – 110 078**

F.15/GGSIPU/Quot./ 2015/794

05.03.2015

Sealed item rate quotations are invited from Manufacture/Distributors/Authorized dealers for the supply of the following technical specifications and terms and conditions:

Name of Equipment	Technical Specification
pH Meter	pH Range: 0.00 to 14.00 pH
	pH Resolution: 0.01pH
	pH Accuracy: $\pm 0.01$ pH
	pH Calibration Points: Up to 3
	Temperature Range: 0 to 100°C
	Temperature Resolution: 0.1°C
	Meter Temperature: ATC/MTC (0 to 100°C)

**Terms and Conditions:**

1. The firms must be manufacture/authorized dealers/authorized distributor.
2. For: GGSIP University, Sector-16C, Dwarka, New Delhi-110078.
3. Payment: After satisfactory installation. No advance payment will be made.
4. Delivery: Within 30 days
5. Taxes: Tax as applicable may please be mentioned in the quotation.
6. Warranty: At least one year
7. Quotation received after due date and time shall be summarily rejected.

Last date, time and venue for submission of quotation	20/03/2015 up to 05:00p.m. at Room No. 306, B-Block, Guru Gobind Singh Indraprastha University, Sector-16C, Dwarka, New Delhi-110078
---	---

Prof. A.K. Narula

For  
5/3/15

Copy to:

1. In-charge, Server Room – with a request to upload the same on University website.
2. Guard file