

University School of Chemical Technology
GURU GOBIND SINGH INDRAPRASTHA UNIVERSITY

Sector: 16C, Dwarka, New Delhi-110 078

Ref. No.: IPU/USCT/Pur/2014/
Date: 22.08.14

Sealed item rate quotations are invited on behalf of Dean, University School of Chemical Technology, Guru Gobind Singh Indraprastha University, Sector:16C, Dwarka, New Delhi-110078 **from manufacturers/distributors/authorized dealers** for the supply of following item with the following technical specifications and terms and conditions:

Sr. No.	Name of the items	Technical Specifications
1.	Analytical Balance (Sartorius)	<p>Model: BSA 224S-CW</p> <p>Weighing Capacity: 220 g</p> <p>Readability: 0.0001 g</p> <p>Pan Size: 90 mm diameter</p> <p>Repeatability: $\leq \pm 0.1$ mg</p> <p>Linearity: $\leq \pm 0.2$ mg</p> <p>Calibration: Internal</p> <p>Feature: Balance should be covered with transparent material, built in motorized calibration weight for automatic internal calibration function via touch key.</p>

Terms & Conditions:

1. The firms must be manufacturer/authorized dealer/authorized distributor.
2. FOR: GGS Indraprastha University, Sector:16C, Dwarka, New Delhi-110078
3. Payment: After satisfactory Installation. No advance payment will be made.
4. Delivery: 30 days
5. Taxes: Tax as applicable may please be clearly mentioned in the quotation.
6. Warranty: One year
7. Quotations received after due date & time shall be summarily rejected.

Last date, time and venue for submission of quotation	05.09. 2014 Up to 5:00 P.M. at Room No. 105, Block B, USCT, GGSIPU, Sector 16 C, Dwarka, New Delhi - 110078
---	--

22/8/2014.