

University School of Chemical Technology
GGS Indraprastha University
Sector 16-C, Dwarka New Delhi
Phone No. 011-25302462

GGSIPIU/USCT/2014-15
February 13, 2015

Sealed item rate quotation are invited on behalf of Dean University School of Chemical Technology, Guru Gobind Singh Indraprastha University, Sector 16C, Dwarka, New Delhi-110078 from Manufacture/ Distributors/ Authorized dealers for the supply of the following technical specification and terms and conditions:

Name of Equipment	Technical Specification
Analytical Balance Model – BSA 224S-CW	Max. Weighing Capacity : 220g Readability : 0.0001g Pan Size : 90mm diameter Repeatability : $\leq \pm 0.1$ mg Linearity : $\leq \pm 0.2$ mg Calibration/Adjustment : Internal Feature: Balance should be covered with transparent material during measurement, built in motorized calibration weight for automatic internal calibration function via touch key.

Terms and Conditions:

1. The firms must be manufacture/authorized dealers/ authorized distributor.
2. For: GGS Indraprastha University, 16C.Dwarka,New Delhi-110078
3. Payment: After satisfactory installation. No advance payment will be made
4. Delivery: 30Days
5. Taxes: Tax as applicable may please be mentioned in the quotation
6. Warranty: At least One year
7. Quotation received after due date and time shall be summarily rejected.
- 8.

Last date, time and venue for submission of quotation	27.02.2015 up to 5PM at Room No. 104,Block-B USCT,GGSIPU.Sector16C, Dwarka, New Delhi -110078
---	--

Please upload.

(Signature)

13/2/2015