



Guru Gobind Singh Indraprastha University  
Sector 16-C, Dwarka, New Delhi-110078

**University School of Biotechnology**

**Corrigendum to the tender notice dated 24<sup>th</sup> December, 2012 uploaded on the GGSIPU website for purchase of -80°C.**

This is with reference to the call for quotation uploaded on GGSIPU website for the purchase of equipments in the DBT project entitled “Circulating miRNA in Esophageal Cancer Patients & Healthy Subjects: Potential blood based Biomarkers” on 24<sup>th</sup> December, 2012.

The corrections are follows:

Sealed item rate quotations are invited on behalf of Registrar, Guru Gobind Singh Indraprastha University, Sector 16 C, Dwarka, New Delhi **from eligible manufacturers/distributors/authorized dealers** for supply of :- **-80<sup>0</sup> C Deep Freezer with stabilizer** as per schedule of quantity attached with this NIQ:-

1.	Name of work:-	<b>Purchase of -80<sup>0</sup> C Deep Freezer with stabilizer</b>
2.	Last date, time and venue for submission of quotation	16th January,2013 Up to 5:00 p.m. at the Room No. 202, Block A, USBT, GGSIPU, Sector 16 C, Dwarka, New Delhi - 110078
3.	Date, time and venue for opening of technical bid	17 <sup>th</sup> October,2012 From 12:00 noon at the Room No. 205, Block A, USBT, GGSIPU, Sector 16 C, Dwarka, New Delhi - 110078
4.	Financial bid shall be opened after evaluation of technical bid/time notified thereafter.	
5.	Earnest Money Deposit	Rs. 13,500 (Rupees Thirteen Thousand five Hundred only)

The bids shall be submitted in two stages viz.(i) *Technical bid* (ii) *Financial bid*.

**Eligibility:-**

- Certificate of Authorized dealership/distributor/manufacturer. (in case of manufacturer, they will self certify so. Authorized dealers/distributor shall attach attested copies of dealership/distributorship certificates issued by the manufacturer)*
- Undertaking by the agency in its Letterhead that:-*
  - that it has not been barred or blacklisted by any of the Central/State Government/Departments/Organizations/Central or State PSU*
  - that it will ensure fair trade practice.*
  - that the proprietor/partners of the agency have any relative employed with GGSIP University.*
- Should have valid registration with DVAT Deptt. of Govt. of Delhi*
- That the item(s) supplied under this contract shall have on-site and comprehensive warranty for 36 months from the date of supply.*

5. *Terms & Conditions:-*

*The bidder shall place his bids in two separate envelopes marked "Technical Bid" and "Financial bid". All documents in support of eligibility as well as another envelop containing DD/Pay order for EMD shall be placed in the envelope marked "Technical Bid". The offered rates shall be placed in the envelop marked "Financial bid". Both these bids should be superscribed with name of work and shall be placed in a third envelop which shall be superscribed "Quotation for -80<sup>0</sup> C Deep Freezer with stabilizer".*

- (1) Bids without EMD will be summarily rejected.*
- (2) Conditional Bids will be summarily rejected.*
- (3) Bids received after due date & time shall be summarily rejected.*
- (4) The "Financial bid" of those bidders whose technical bids have qualified will only be opened.*
- (5) Rate must be quoted in Indian Rupees only net in figures & words inclusive of taxes, levies, cartage handling, loading, unloading etc.*
- (6) Delivery :- F.O.R. GGSIP University, Sector 16 C, Dwarka, New Delhi*
- (7) University will provide the custom duty exemption certificate.*
- (8) Delivery period:- 42 days from the date of supply order.*
- (9) The EMD of unsuccessful bidders shall be refunded immediately.*
- (10) The successful bidder have to submit a Performance Security Deposit @ 5% of the quoted value in the form of Demand Draft/Pay order drawn in favour of Registrar, GGSIPU, Delhi within 7(seven) days of the communication accepting the bid. EMD shall be adjusted toward Performance Security Deposit. The Performance Security Deposit shall be refunded without interest after completion of the guarantee period of 15 months.*
- (11) In case the successful bidder fails to deposit the Performance Security within the stipulated 7 (seven) days of the communication accepting the bid, the EMD shall be forfeited to GGSIP University absolutely.*
- (12) In case the successful bidder fails to supply the item(s) within the delivered period, a sum equal to 0.5% of the contract price per week or part thereof until the actual delivery subject to maximum of 10% of the value of supply order shall be deducted.*
- (13) The payment will be made after delivery and satisfactory installation of the equipment.*
- (14) Taxes etc., if any, leviable shall be deducted at source.*
- (15) The validity of the bid will be 120 days from the date of opening of financial bids. During the validity period, the successful bidder shall not be allowed to withdraw. In case of withdrawal, the EMD shall be forfeited to GGSIP University absolutely and no claim shall be admitted in this regard. Such bidder shall not be allowed to participate in the re-quotationing process.*
- (16) The rates of successful bidder will be valid for 12 months from the date of issue of letter of acceptance.*
- (17) University reserves the right to reject any or all the bids or accept them in part or reject the lowest bid without assigning any reason.*
- (18) Unauthorized substitution or materials delivered in error of wrong description or quality or supplied in excess quantity or rejected goods shall be returned to the bidder at bidder's cost & risk.*
- (19) The successful bidder shall make all arrangements towards safe and complete delivery at the designated location indicated in the supply order. Such responsibility on the part of the bidder will include taking care of insurance, freight, state level permits etc. as applicable.*
- (20) In case of any dispute relating to meaning, scope, manufacturing, operation or effect of this contract or the validity or the breach thereof, University and the contractor shall make every effort to resolve amicably by direct discussion/negotiation.*
- (21) In case the dispute cannot be settled amicably within 30 days of the raising of dispute by either party, either party may seek settlement of the dispute by arbitration in accordance with the provisions of the Arbitration & Conciliation Act, 1996 and the award made in pursuance thereof shall be binding on all the parties. The sole arbitrator shall be appointed by Vice Chancellor, GGS Indraprastha University.*

- (22) *The performance under this contract shall not be stopped for any reason whatsoever during the said dispute/proceedings unless the contractor is specifically directed to do so by the University.*
- (23) *The venue of arbitration proceedings shall be Delhi/New Delhi. The language of proceedings shall be English. The law governing the substantive issues between the parties shall be the Laws of India. All disputes are subject to Jurisdiction of Delhi Courts only.*
- (24) *It is also a term of the contract that if any fee payable to the arbitrator, shall be paid equally by both the parties. It is also a term of the contract that the arbitrator shall be deemed to have entered in the reference on the date he/she issues notice to both the parties calling them to submit their statement of claims and counter statement of claims.*
- (25) *Force Majeure.*  
*For purpose of this clause, 'Force Majeure' means an event beyond the control of the contractor and not involving the contractor's fault or negligence and not foreseeable. Such events may include, but are not limited to, acts of the University either in its sovereign or contractual capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargo.*  
*If a Force Majeure situation arises, the contractor shall promptly notify the University in writing of such conditions and cause thereof. Unless otherwise directed by the University in writing, the contractor shall continue to perform its obligations under this contract as far as reasonably practical and shall seek all reasonable alternative means for performance not prevented by Force Majeure event.*

This NIT has also been uploaded on University website ([www.ipu.ac.in](http://www.ipu.ac.in))

### SCHEDULE OF QUANTITY

Name of work:- **-80<sup>0</sup> C Deep Freezer with stabilizer**

S.No	Particulars/item	Qty.	Unit	Rate	
				Amount in fig.	Amount in words
1	<p><b>TYPE : Upright</b></p> <p>INTERNAL : 1265H x 550W x DIMENSION 588D (mm)</p> <p>EXTERNAL : 1930H x 800W x DIMENSION 855D (mm)</p> <p>CAPACITY : 400 - 450 Liters.</p> <p>OPERATING : -50°C to -86°C TEMPERTURE with 1°C increments, even at ambient temp up to 32 °C .</p> <p>POWER : Less than CONSUMPTION 13kWh/Day</p> <p>ELECTRIC : 220V/50Hz, 13 SUPPLY Amps. single phase</p> <p>✓ Fully programmable microprocessor controlled with membrane keypad and eye level control panel.</p> <p>✓ <b>System should have 304L grade stainless steel interior and tough, powder coated</b></p>	1			

	<p>exterior finish constructed on steel gauge of 18g ZINTEC 1.2 min.</p> <ul style="list-style-type: none"> <li>✓ <b>Freezer should have five insulated doors giving access to shelves which should be adjustable in 13mm steps, must come with 4 adjustable shelves and make the floor cabinet total five inner storage compartments.</b></li> <li>✓ <b>Freezers should have heated air vent and front panel air filter.</b></li> <li>✓ Heavy duty lockable castors and lockable outer doors and lids.</li> <li>✓ <b>Freezer must have battery back-up and 4 PIN security lock for unauthorized tempering.</b></li> <li>✓ Freezer must have RS 232 interface data logging port and it must also have on board SMART diagnostic software.</li> <li>✓ Audible and visible alarms for temperature, power failure, system failure, battery low etc. and it also have remote alarm port for connection to and auto dialer.</li> <li>✓ Freezer must use CFC-FREE, HCFC-FREE non flammable refrigerants, and refrigeration system must be energy efficient and hermetically sealed cascade refrigeration system. <b>Compressor should be capable to run any voltage between 190 – 270V. Freezer must have ISO 9001- safety requirements and IEC 61010 Electrical safeties CE certified.</b></li> <li>✓ Freezer must have washable condenser filter indication which should keep fins free of dust to maintain peak cooling efficiency.</li> <li>✓ <b>A suitable voltage stabilizer to be provided by the supplier.</b></li> </ul>			
--	---	--	--	--

Name, signature & designation of issuing authority