University School of Chemical Technology GGS Indraprastha University Sector 16-C, Dawarka New Delhi Phone No. 011-25302462

uplera

GGSIPU/USCT/2015 /21 89/L January 28, 2015

Sealed item rate quotation are invited on behalf of Dean University School of Chemical Technology, Guru Gobind Singh Indraprastha University, Sector 16C, Dwarka, New Delhi-110078 from Manufacture/ Distributors/ Authorized dealers for the supply of the following technical specification and terms and conditions:

Name of Equipment	Technical Specification
Ultra Sonicator	Electric Supply: 170-245v, A>C, 50Hz 1 phase.
Bath	Ultrasonic Power: 40kHz or 33± 3kHz
1 Tank size: 300×150×150mm	Ultrasonic Generator: Micro Controller base Digital Timer.
2 Basket size: 270×100×100mn	Htraconianton, Wish Land MODERT
3 Capacity: 6.5litre	Degassing facility.
	Automatic frequency tunings for maximum output.
	Modular generator for easier modification/servicing
,	Timer: Two digit LED display 0-20/90 minutes timer with Settable
İ	& On/Off keys.
	Auto degassing: Preset 10min., to remove dissolved gases from solution to increase Ultrasonic efficiency.
	Temperature control: By Digital temp. controller with digit LED
	Display, up/ Down keys for setting.
ii	Settable temperature: ambient temperature to 60°c
	Tank MOC: S.S 304 grade, 1.6mm thick & buffed from inside
*	Derain Valve: ½" ball valve is provided for cleaning the tank
	Accessory: Basket, lid, manual is provided
	,

## Terms and Conditions:

- The firms must be manufacture/authorized dealers/ authorized distributor.
- For: GGS Indraprastha University, 16C, Dwarka, New Delhi-110078
- Payment: After satisfactory installation. No advance payment will be made
- Delivery: 30Days
- Taxes: Tax as applicable may please be mentioned in the quotation
- Warranty: At least One year
- Quotation received after due date and time shall be summarily rejected.

Last date, time and venue for submission of quotation US	3.02.2015 up to 5PM at Room No. 104,Block-B SCT,GGSIPU,Sector16C, Dwarka, New Delhi -110078
--	--