

web. upload

University School of Chemical Technology  
GGS Indraprastha University  
Sector 16-C, Dwarka New Delhi  
Phone No. 011-25302462

GGSIPU/USCT/2014/2087/L  
December , 2014


\* Sealed item rate quotation are invited on behalf of Dean University School of Chemical Technology, Guru Gobind Singh Indraprastha University, Sector 16C, Dwarka, New Delhi-110078 from Manufacture/ Distributors/ Authorized dealers for the supply of the following technical specification and terms and conditions:

Name of Equipment	Technical Specification
Sartorius Precision Balance Model – BSA 822S-CW	Max. Weighing Capacity : 820g Readability : 0.01mg Pan Size : 150mm Repeatability : $\leq \pm 0.02\text{mg}$ Linearity : $\leq \pm 0.03\text{mg}$ Calibration/Adjustment : Internal Calibration with auto Adjustment External Calibration with Auto Adjustment

Terms and Conditions:

1. The firms must be manufacture/authorized dealers/ authorized distributor.
2. For: GGS Indraprastha University, 16C, Dwarka, New Delhi-110078
3. Payment: After satisfactory installation. No advance payment will be made
4. Delivery: 30Days
5. Taxes: Tax as applicable may please be mentioned in the quotation
6. Warranty: At least One year
7. Quotation received after due date and time shall be summarily rejected.
- 8.

Last date, time and venue for submission of quotation	09.01.2015 up to 5PM at Room No. 104, Block-B USCT, GGSIPU, Sector 16C, Dwarka, New Delhi -110078
---	--

  
19/12/2014